

**Propulse Ear
Irrigator**
Maintenance &
Safety
Inspections

To ensure optimum performance the Propulse Ear Irrigator should be serviced every 12 months. Service or repairs conducted by unauthorised agencies/organisations invalidate any or implied warranties from Mirage.

The Propulse Ear Irrigator should under-go routine electrical safety testing to ensure that it remains safe to use.

Users of the Propulse Ear Irrigator should carry out regular inspections to ensure that the handle and hose, power adaptor and cable, reservoir and main body of the unit are free from damage prior to use. If any damage is evident the Propulse Ear Irrigator should NOT be used until replacement parts have been fitted. Only Propulse branded items should be used with the Propulse Ear Irrigator.

The Propulse Ear Irrigator is not user repairable and should be returned to your Propulse supplier or Mirage Service Centre (UK only) for service/and or repair:

Mirage Service Centre
1 Little Mundells
Welwyn Garden City
Hertfordshire
AL7 1EW
UK

mirage[®]
Health Group

Mirage Health Group Ltd.

1 Little Mundells,
Welwyn Garden City,
Hertfordshire
AL7 1EW

UK

Tel: +44 (0) 845 130 5440

Fax: +44 (0) 845 130 6440